

1. Information of the substance/preparation and of the company	
1.1 Commercial product name	CK-MB Calibrator
Cat. No.	KG2393
1.2. Company	<b>BGT BioGenTechnologies GmbH</b> Von-Langen-Weg 10 D-48565 Steinfurt Tel.: 02551/4090 Fax.: 02551/1298
1.3. In Emergencies	Tel.: 02551/4090

2. Composition / Hazardous Ingredients	
	Contains no hazardous substances in reportable quantities.
Classification:	No hazardous chemicals as classified by CHIP 3

3. Hazards identification	
Warnings:	For <i>in vitro</i> diagnostic use only. Do not pipette by mouth. Handle laboratory reagents in accordance with Good Laboratory Practice. Biohazard - The material in this kit is of human origin. This material has been tested for the HIV (Human Immunodeficiency Virus) Antibody, HBs Ag and HCV and found to be non-reactive. However as no method can offer complete assurance as to the absence of infectious agents. This material and all patient samples should be handled as though capable of transmitting infectious disease and disposed of accordingly.
Risk phrases:	None
Safety phrases:	None

4. First aid measures	
Emergency First Aid:	Call an internal First Aider where available or else contact a physician immediately.  Direction for First Aider or Physician:
Inhalation:	If inhaled, remove to fresh air. If breathing becomes difficult, call a physician.
Ingestion:	If swallowed wash out mouth with water provided person is conscious, call a physician.
Skin Contact:	In case of skin contact immediately wash skin with soap and copious quantities of water.
Eye Contact:	Flush with copious amounts of water for at least 15 minutes. Assure adequate flushing by separating eyelids with fingers. Call a physician.

Commercial product name:	CK-MB Calibrator
--------------------------	------------------

<b>5. Fire-fighting measures</b>	
Flash point:	Not determined
Flammable limits:	Not determined
Extinguishing media:	No restriction
Special Fire Fighting Procedures:	Wear self contained breathing apparatus and protective clothing to prevent contact with skin and eyes.
Unusual Fire and Explosion Hazards:	May emit toxic fumes under fire conditions
Thermal decomposition:	Not determined

<b>6. Accidental release measures</b>	
After spillage:	<p><b>PRECAUTIONS</b> Treat material as you would human body fluids. Disposal should, therefore include autoclaving at 121°C for 20 mins prior to appropriate disposal with regard to local regulations.</p> <p>Alternatively, decontamination using a 0.1% sodium hypochlorite is acceptable. Perform this type of decontamination in a fume hood.</p>
Absorbent material:	No restriction

<b>7. Handling and storage</b>	
7.1 Handling:	Wash thoroughly after handling. Do not use if skin is cut or scratched. Avoid contact with eyes, skin and clothing.
7.2 Storage	Store at temperatures and conditions as indicated on the product label
Do not store together:	N/A

<b>8. Exposure controls/personal protection</b>	
MEL/OES	Not determined
Respiratory protection:	None
Eye protection:	Protective glasses
Hand protection:	Standard laboratory rubber or latex gloves.
Skin protection:	A laboratory coat or apron is recommended

<b>Commercial product name:</b>	CK-MB Calibrator
---------------------------------	------------------

<b>9. Physical and chemical properties</b>	
Physical state:	Lyophilised
Colour:	Pale yellow
Odour:	Odourless
pH value:	Not determined
Boiling point	Not determined
Melting point:	Not determined
Oxidising properties:	Not determined
Ignition temperature	Not determined
Explosion limits:	Not determined
Density:	Not determined
Solubility in water:	Soluble

<b>10. Stability and reactivity</b>	
Stability:	Stable
Conditions to avoid:	Not determined
Substances to avoid:	Strong oxidising agents, also acids.
Hazardous decomposition products:	Not determined

<b>11. Toxicological information:</b>	
LD50 oral	Not determined
LD50 intraperitoneal	Not determined
LD50 skin	Not determined
LC50 inhalation	Not determined
Toxicologic information:	Not thoroughly investigated

<b>12. Ecological information</b>	
Water hazard class:	Data not yet available

<b>Commercial product name:</b>	CK-MB Calibrator
---------------------------------	------------------

<b>13. Disposal</b>	<p><b>PRECAUTIONS</b> Treat material as you would human body fluids. Disposal should, therefore include autoclaving at 121°C for 20 mins prior to appropriate disposal with regard to local regulations.</p> <p>Alternatively, decontamination using a 0.1% sodium hypochlorite is acceptable. Perform this type of decontamination in a fume hood.</p>
---------------------	---

<b>14. Transport Information</b>			
GGVE/GGVS:	Data not yet available	RID/ADR	Data not yet available
ICAO/IATA-DGR:	Data not yet available	UN-Nr.:	Data not yet available
IMDG-CODE:	Data not yet available	MFAG	Data not yet available
		EMS:	Data not yet available

<b>15. Regulatory information</b>	
	This product contains no hazardous chemicals in reportable quantities according to CHIP 3
Hazard symbol:	None
Risk phrases:	None
Safety phrases:	None
Precautions:	See Section 3

<b>16. Other information</b>	
	The information herein is believed to be correct as of the date hereof but is provided without warranty of any kind. The recipient of our products is responsible for observing any National laws and guidelines applicable.