

1. Information of the substance/preparation and of the company	
1.1 Commercial product name	Liquid Protein Calibrators
Cat. No.	KG2692
1.2. Company	BGT BioGenTechnologies GmbH Von-Langen-Weg 10 D-48565 Steinfurt Tel.: 02551/4090 Fax.: 02551/1298
1.3. In Emergencies	Tel.: 02551/4090

2. Composition / Hazardous Ingredients	
Chemical Name	(Contains no hazardous substances in reportable quantities)
Classification:	No hazardous chemicals as classified by CHIP 3

3. Hazards identification	
Warnings:	<p>For <i>in vitro</i> diagnostic use only. Do not pipette by mouth. Handle laboratory reagents in accordance with Good Laboratory Practice.</p> <p>Calibrators contain sodium azide (<0.1%w/v). May cause irritation to skin and eyes, may be irritating to mucous membranes and upper respiratory tract. Avoid contact with skin, avoid inhalation and ingestion.</p> <p>Sodium azide reacts with lead or copper plumbing to form potentially explosive azides. When disposing of such reagents flush with large volumes of water to prevent azide build up. Exposed metal surfaces should be cleaned with 10% sodium hydroxide.</p> <p>As no test method can offer complete assurance that products derived from human source will not transmit infectious agents, it is recommended that this product be handled with the same precautions used for patient samples.</p>
Risk phrases:	None
Safety phrases:	None

4. First aid measures	
Emergency First Aid:	Call an internal First Aider where available or else contact a physician immediately. Direction for First Aider or Physician:
Inhalation:	If inhaled, remove to fresh air. If breathing becomes difficult, call a physician.
Ingestion:	If swallowed wash out mouth with water provided person is conscious, call a physician.
Skin Contact:	Wash off skin thoroughly with water. Remove contaminated clothing and treat with dilute Virkon [®] (1%) briefly. Wash skin in detergent and water after swabbing with Virkon [®] or similar Virucidal disinfectant.
Eye Contact:	Flush with copious amounts of water for at least 15 minutes. Assure adequate flushing by separating eyelids with fingers. Call a physician.

Commercial product name:	Liquid Protein Calibrators
---------------------------------	----------------------------

5. Fire-fighting measures	
Flash point:	Not determined
Flammable limits:	Not determined
Extinguishing media:	No restriction
Special Fire Fighting Procedures:	Wear self-contained breathing apparatus and protective clothing to prevent contact with skin and eyes.
Unusual Fire and Explosion Hazards:	May emit toxic fumes under fire conditions
Thermal decomposition:	Not determined

6. Accidental release measures	
After spillage:	<p>Use appropriate spill absorbent kit as instructed by the manufacturer. Alternatively, mop up with an absorbent material and hold for waste disposal.</p> <p>Appropriate Personal Protective Equipment should be worn, e.g. laboratory coat, gloves, safety glasses and mask.</p> <p>Ventilate area and wash spill site after material pick up is complete.</p>
Absorbent material:	No restriction

7. Handling and storage	
7.1 Handling:	Wash thoroughly after handling. Do not use if skin is cut or scratched. Avoid contact with eyes, skin and clothing.
7.2 Storage	Store at temperatures and conditions as indicated on the product label
Do not store together:	N/A

8. Exposure controls/personal protection	
MEL/OES	Not determined
Respiratory protection:	None
Eye protection:	Protective glasses
Hand protection:	Standard laboratory rubber or latex gloves.
Skin protection:	A laboratory coat or apron is recommended

Revision: C
Date: 14th May 2003

Commercial product name:	Liquid Protein Calibrators
---------------------------------	----------------------------

9. Physical and chemical properties	
Physical state:	Liquid
Colour:	Colourless to pale straw
Odour:	Not determined
pH value:	Not determined
Boiling point	Not determined
Melting point:	Not determined
Oxidising properties:	Not determined
Ignition temperature	Not determined
Explosion limits:	Not determined
Density:	Not determined
Solubility in water:	Not determined

10. Stability and reactivity	
Stability:	Stable
Conditions to avoid:	Not determined
Substances to avoid:	Strong oxidising agents
Hazardous decomposition products:	Not determined

11. Toxicological information:	
LD50 oral	Not determined
LD50 intraperitoneal	Not determined
LD50 skin	Not determined
LC50 inhalation	Not determined
Toxicologic information:	Not thoroughly investigated

12. Ecological information	
Water hazard class:	Data not yet available

Revision: C
Date: 14th May 2003

Commercial product name:	Liquid Protein Calibrators
---------------------------------	----------------------------

13. Disposal	<p>Treat material as you would human body fluids. Disposal should, therefore include autoclaving at 121°C for 20 mins prior to appropriate disposal with regard to local regulations.</p> <p>Alternatively, decontamination using a 0.1% sodium hypochlorite is acceptable. Perform this type of decontamination in a fume hood.</p>
---------------------	--

14. Transport Information			
GGVE/GGVS:	Data not yet available	RID/ADR	Data not yet available
ICAO/IATA-DGR:	Data not yet available	UN-Nr.:	Data not yet available
IMDG-CODE:	Data not yet available	MFAG	Data not yet available
		EMS:	Data not yet available

15. Regulatory information	
	This product contains no hazardous chemicals in reportable quantities according to CHIP 3.
Hazard symbol:	None
Risk phrases:	None
Safety phrases:	None
Precautions:	Contains human material; handle with due care.

16. Other information	
	The information herein is believed to be correct as of the date hereof but is provided without warranty of any kind. The recipient of our products is responsible for observing any National laws and guidelines applicable.

Revision: C
Date: 14th May 2003